



## PS-2063

### Product Description:


P-DERM® PS-2063 is a 0.19mm thickness perforated silicone gel/acrylic trilaminate consisting of a high initial tack silicone gel skin contact adhesive, polyurethane and medical pressure sensitive acrylic. PS-2063 is supplied with a perforated 75 micron polycarbonate liner on the silicone gel adhesive side and a 125 micron siliconized white polypropylene liner on the acrylic side of the construction.

### Construction / Properties:

Property	Value	Test Method
Total Thickness	0.19mm	QSP-726
Silicone Gel Coat Weight	125 gsm	QSP-724
Polyurethane Carrier	28 micron	QSP-724
Acrylic Adhesive Coat Weight	37 gsm	QSP-724
Silicone Gel Adhesion	0.9 N/25mm	QSP-723
Acrylic Adhesion	3.5 N/25mm	QSP-722
Hole Size / Open Area	1.5 mm / 17%	--
Continuous Use Conditions	10 – 70 C	WI-8.2-13
Biocompatibility	Cytotoxicity: Pass Primary Dermal: Pass Buehler Sensitization: Pass	ISO 10993
Recommended Sterilization Method	Ethylene Oxide	--

### Features

- Excellent Instantaneous Tack
- Atraumatic Removal from Skin
- Hypoallergenic
- Ability to be Removed and Repositioned
- Excellent Converting Properties



Specific tests should be performed by the end user to determine the product suitability for the particular application.

For Additional Information:

E-Mail: [sales@polymerscience.com](mailto:sales@polymerscience.com)

Toll Free: +1 888.533.7004

Web: [www.polymerscience.com](http://www.polymerscience.com)

Revision: 061015